

CERTIFICATE OF ANALYSIS

CUSTOMER

Advanced Hygienic Contracting



CERTIFICATE NO.

BIOCOTE REF.

10/179

SAMPLE DETAILS

DATE RECEIVED

May 2010

Advanced Hygienic Contracting BioCote treated

METHOD: Determination of Antibacterial Activity using Test Based on ISO 22196

DATE ANALYSED

19/05/2010

DATE REPORTED

24/05/2010

RESULTS (as cfu cm²)

SAMPLE	SPECIES	REDUCTION (%)
Treated PVC Plaque	E.coli	99.99%
Treat PVC Plaque	MRSA	>99.94%

The above data shows the difference in the population following contact with the surface of the samples listed for 24 hours at 35°C under a RH of 95% relative to initial population.

BioCote Ltd
Technology Centre
Wolverhampton Science Park
Glaisher Drive
Wolverhampton
WV10 9RU
UK

Technical Manager
Lesley Taylor

A handwritten signature in black ink, appearing to read "Lesley Taylor", is positioned below the printed name.